Certificate No: LR24131651TA
Issue Date: 13/06/2024
Expiry Date: 12/06/2029

Type Approval Certificate

This is to certify that the undernoted product(s) has/have been tested with satisfactory results in accordance with the relevant requirements of the Lloyd's Register Type Approval System.

Manufacturer	**Shanghai Donjun Automation Equipment Co., Ltd**
Address	No.1578, Zhaogao Road, Pudong New Area, Shanghai, 201208, China
Type	Sensor Unit
Description	Portable Oil/Water interface Detector
Trade Name	DJC-U
Application	Marine applications for determination of oil/water interface
Specified Standard	IMO Resolution MEPC.5 (XIII) Specification for oil/water interface detectors
Ratings	Vertical tape length: 15 m or 30 m Accuracy: ±5mm Response time: <3 s Ex marking: Ex ia II C T4 Ga Degree of protection: IP:56 (sensor: IP 68, 0.3Mpa)
Other Conditions	The specific conditions of use for explosion protection shall be complied with as indicated in certificate GYB23.1919X remarks.

Downloaded Copy

Ning Guan

Electrical & Control - Senior Specialist to Lloyd's Register Classification Society (China) Co., Ltd.

12th Floor, Chong Hing Finance Centre, No. 288
Nanjing West Road, Shanghai, 200003, China

Lloyd's Register Group Limited, its affiliates and subsidiaries and their respective officers, employees or agents are, individually and collectively, referred to in this clause as 'Lloyd's Register'. Lloyd's Register assumes no responsibility and shall not be liable to any person for any loss, damage or expense caused by reliance on the information or advice in this document or howsoever provided, unless that person has signed a contract with the relevant Lloyd's Register entity for the provision of this information or advice and in that case any responsibility or liability is exclusively on the terms and conditions set out in that contract.

Certificate No: LR24131651TA
Issue Date: 13/06/2024
Expiry Date: 12/06/2029

Type Approval Certificate

This certificate is not valid for equipment, the design, ratings or operating parameters of which have been varied from the specimen tested. The manufacturer should notify Lloyd's Register Classification Society (China) Co., Ltd. of any modification or changes to the equipment in order to obtain a valid Certificate.

Previous Version: **19/80026**

The Design Appraisal Document LR24131651TA and its supplementary Type Approval Terms and Conditions form part of this Certificate.

Downloaded Copy

12th Floor, Chong Hing Finance Centre, No. 288
Nanjing West Road, Shanghai, 200003, China

Lloyd's Register Group Limited, its affiliates and subsidiaries and their respective officers, employees or agents are, individually and collectively, referred to in this clause as 'Lloyd's Register'. Lloyd's Register assumes no responsibility and shall not be liable to any person for any loss, damage or expense caused by reliance on the information or advice in this document or howsoever provided, unless that person has signed a contract with the relevant Lloyd's Register entity for the provision of this information or advice and in that case any responsibility or liability is exclusively on the terms and conditions set out in that contract.